



EDINBURGH ASSAY OFFICE

Tom Anderson
Hobart House
80 Hanover Street
Edinburgh
EH1 3QN

Friday, 20 January 2023

1 x copper sheet

Our Ref: 40421720A

Scope

The Assay Office was asked to examine the above item and comment on the following:

- List all elements present in the alloy



Copper sheet

Analysis and Inspection

The Assay Office have analysed the items by XRF X-ray spectroscopy. Our findings are as follows:

- The sample tested contain 99.99% copper.

Conclusion

If you require a result carried out for trace elements, a destructive chemical analysis by ICP must be carried out.

Please note that whilst we are UKAS accredited for XRF testing, this particular report is out with the scope of accreditation.

I trust the above information is satisfactory, but please do not hesitate to contact me should you require further information or clarification.

Yours sincerely,

Javier Quesada Perez
Senior Laboratory Analyst

PAGE 1 OF 1